

Avery Weigh-Tronix

Calibration Certificate

Avery Weigh-Tronix
 Foundry Lane, Smethwick
 West Midlands, B66 2LP
www.averyweigh-tronix.com

Tel: 0845 900 22 44



Customer Details	
Name :	Chas Storer Ltd
Address :	Coopers Lane Northaw Potters Bar EN6 4NE
Contact Name :	John Curtis
Telephone No.	01707656261

Linearity / Accuracy Test				
Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0 kg			0 kg	0 kg
4,000 kg			4,000 kg	0 kg
8,000 kg			8,000 kg	0 kg
12,000 kg			12,000 kg	0 kg
16,000 kg			16,000 kg	0 kg
24,800 kg			24,800 kg	0 kg
40,800 kg			40,800 kg	0 kg
46,720 kg			46,720 kg	0 kg

Certificate No. 231029395/0101
Date of Calibration : 04/07/2018
Next Calibration Due : Jul 2019

Agreement No.	3041000001
Location :	POTTERS BAR
Department :	
Plant Number :	
Manufacturer :	Avery Weigh-Tronix
Model :	E1205
Serial No :	151250740
Capacity :	50000 kg
Resolution :	20kg
Tolerance :	+/-60kg
Calibration Frequency :	

Test Point	Off Centre 2 3 Loading Test 1 4		End -Middle-End 1-2-3	
	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
9,940 kg			9,940 kg	0 kg
9,940 kg			9,940 kg	0 kg
9,940 kg			9,940 kg	0 kg
9,940 kg			9,940 kg	0 kg

Test Weight Details
Weight Set No. & Certificate No.
W012: A0282
Test weights and equipment are traceable to national standards.

Repeatability Test				
Test Point	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
40,800 kg			40,800 kg	0 kg
40,800 kg			40,800 kg	0 kg
40,800 kg			40,800 kg	0 kg

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments.

Comments Machine not tested to full capacity Make-up weight used No as found readings: Machine repaired	Tolerances Unless otherwise agreed tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows: 0-500 divisions +/- 1 divn. >500-2000 divisions +/- 2 divn. >2000 divisions +/- 3 divn.
	For key calibration information please login to https://averyweigh-tronix.com

Technicians Name: Wainika M

Customer Name: Craig Curtis

Technicians Signature: *W.P.L.*

Customer Signature: *C.C.*